



क्रमांक: आईसीएमआर-निरेह/बीपीएल/PUR/2019-20/VD/114/1625

दिनांक: 31.01.2020

QUOTATION/ RATE INQUIRY

Sealed quotations are hereby invited on behalf of Director, ICMR-National Institute for Research in Environmental Health from the Aspirant Manufacturer/ Authorized Dealer/ Distributor for 3 (three) No. of “Noise Dosimeter” in the prescribed format attached to this mail. Specification of the Noise Dosimeter is appended below:

| Particulars | Specifications |
|--|--|
| Standards | IEC/ISO and ANSI |
| Measurement parameters | Sound Exposure & Estimated Sound Exposure, % Dose & Estimated % Dose, ULT Duration, SPL Max Level & Time, SPL Min Level Dose, Identity Field, L(C-A), Lavg / Leq, Max, Min, Peak, Projected Dose, Run Time, SEL (Lep), SPL, Threshold, TWA, Upper Limit (UL) |
| Measurement range | RMS Range - 60 dB(A) to 140 dB(A) Peak Range - 80 dB(C) to 143 dB(C) Octave Bands - 70 dB(A) to 140 dB(A) |
| Time Weighting | Fast, Slow, Impulse |
| Frequency Weighting | A, C or Z |
| Octave Band Parameters | 1:1 analysis, 20Hz – 10kHz |
| Exchange rate | 3, 4 or 5 dB |
| Calibration | Automated |
| Data Logging | 1 second sample |
| Memory | Minimum 72 hours of data |
| Run time | Minimum 36 hours - Broadband Mode Minimum 16 hours - Octave mode |
| Maximum run duration of any single measurement | 24 hours |
| Audio recording | Audio events recording on trigger, up to 100 hours of recording |
| Battery | Rechargeable |
| Environmental conditions | -10 °C to 50°C |
| Operating humidity | Up to 90% Relative Humidity, non-condensing |
| Weight | < 120 gm (with batteries) |
| Warranty | Preferably for 3 years |

Please submit duly sealed quotations, separately to the Office of the Director, ICMR-NIREH, Bhopal Bypass Road, Bhauri, Bhopal - 462030 latest by 21.02.2020 up to 11:00 AM.

Please note the following:-

1. Open Envelope and unsigned Quotations will be rejected.
2. Quotations send by E-mail / Fax will not be considered.
3. Quotations to be submitted in the attached prescribed proforma only in a sealed cover.
4. Valid authorization certificate from the relevant company must be attached to the quoted items. Failing to which quotation will be rejected.
5. Quotation received after the last date of submission is liable to be rejected.
6. Authorized dealer can submit quotation on behalf of only one principal/ OEM.
7. GST rates applicable on public funded research institutes have been revised to 05%.
8. Director, NIREH reserves the right to reject any or all quotations without assigning any reason.
9. Please clearly mark "Quotation No. "ICMR-NIREH/BPL/PUR/2019-20/VD/114/1625" and Due Date: 21.02.2020 on top of the envelope.

Sd/-xxx
Incharge Purchase
for Director

To,
Director,
ICMR-NIREH,
Bhopal

Sub: Offer for quotation no. ICMR-NIREH/BPL/PUR/2019- 20/VD/114/1625

With reference to above, our offer is as under:-

| S No | Particulars | Req Qty | Company/ Brand Name | Unit Price | Discount in % | GST @ 5% | Price to NIREH (Including GST & Discount) | Net Price (Qty * Price to NIREH) |
|------|-------------|---------|------------------------|------------|---------------|----------|---|----------------------------------|
| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 |
| | | 3 | | | | | | |

I/We, the undersigned, hereby declare that:

1. I/We certify that quoted product meets the specification as described in the call for quotation.
2. I/ We certify that rate quoted for this product does not exceed the Maximum Retail Price of the product.

Name and Surname:
Signature of Bidder:
Stamp of the firm/company